AHMEDABAD MUNICIPAL CORPORATION
CENTRAL MEDICAL STORE
T.Notice No -8/CMS/AMC/2008-09

Tenders are invited for the purchase of following items for the use of S.C.L.Hospital, L.G.Hospital and C.H.Nagari Hospital, Ahmedabad on rate contract (one year) and quantity basis from date of sanction.

Hospital Equipments			
Tender No.	**Item Name**	**Appx.quantity**	**E.M.D.**
T 1	Defibrillator	20	Rs.25000
T 2	Syringe Infusion Pump	6	Rs.3000
T 3	Pulse Oxymeter with ECG Monitor	4	Rs.3000
T 4	Pulse Oxymeter	2	Rs.1000
T 5	Five function Multipara Monitor	2	Rs.3000
T 6	Capnograph with pulse Oxymeter	4	Rs.5000
T 7	Multipara Monitor	6	Rs.5000
T 8	Vital sign Monitor	1	Rs.5000
T 9	Ventilator (Adult)	2	Rs.3000
T 10	T.M.T. Machine (Computerized stress Test system)	1	Rs.15000
T 11	Twin Dome ceiling suspended surgical O.T.Light	8	Rs.25000
T 12	Sonic Ultrasound scalpel (knife)	1	Rs.15000
T 13	Radio lucent operation Table	1	Rs.5000
T 14	'0' Degree Adult size Telescope for Cytoscope	1	Rs.3000
T 15	Laproscopic three chip camera	1	Rs.5000
T 16	Digital Video colposcope	1	Rs.5000
T 17	Thermachoice Uterine Balloon Therapy system	1	Rs.10000
T 18	Laproscopic Instruments	1 SET	Rs.10000
T 19	X-Ray Machine 500 m. A.	3	Rs.20000
T 20	Arthroscopic Light source	1	Rs.3000
T 21	Short wave Diathermy (400 w)	1	Rs.1000
T 22	Continuous passive Motion (CPM)	1	Rs.1000
T 23	Stimulator or Combination Therapy Unit	1	Rs.1000
T 24	Scanner Laser	1	Rs.3000
T 25	EMG / NCV Evoked potential windows based system (Electo Myography & Nerve condition)	1	Rs.3000

T 26	Co2 Laser with Pattern Generator	1	Rs.10000
T 27	Skin Bank Refrigerator (Deepfreeze)	1	Rs.3000
T 28	Microsurgical Instruments [Ergonomic Design]	1SET	Rs.3000
T 29	Surgical Operating Microscope for Plastic surgery	1	Rs.25000
T 30	Haemo Mixture	2	Rs.1000
T 31	Blood collection chair	2	Rs.1000
T 32	Blood collection Monitor	1	Rs.1000
T 33	Di-Electric Tube sealer (Table Top Model)	1	Rs.1000
T 34	Elisa Reader	1	Rs.3000
T 35	Semi Automated Micro plate washer	1	Rs.3000
T 36	Semi Auto Biochemistry Analyzer (Backup Machine)	1	Rs.3000
T 37	Blood Bank Refrigerator	1	Rs.3000
T 38	Hot Plate	1	Rs.1000
T 39	Phaco Machine	1	Rs.10000
T 40	Auto Retrectometer	1	Rs.3000
T 41	Slit Lamp	1	Rs.1000
T 42	A- Scan Machine	1	Rs.10000
T 43	Impendence Audiometer	1	Rs.3000
T 44	Laser Lenses	4	Rs.5000
T 45	Vitrectomy Lenses	5	Rs.10000
T 46	Horizontal Sterilizer	2	Rs.10000
T 47	Indirect Ophthalmoscope with CCTV [Binocular]	1	Rs.5000
T 48	ERG Machine	1	Rs.20000
T 49	HRT-III	1	Rs.20000
T 50	Vitrectomy Machine	1	Rs.25000
T 51	Liquid Medical Oxygen Supply System	2	Rs.5000
T 52	Portable Sonography Machine With Nerve Locator	1	Rs.15000
T 53	Fiber Optic Flexible Intubating Bronchoscope	1	Rs.5000

Only manufacturer or authorized distributor with valid authority may apply. Manufacturer or authorized distributor if interested may download the tender document from AMC website www.egovamc.com & submit as per schedule given below.

Tender fee for **each tender** is Rs.500 (non refundable) and EMD as given in table for each tender each item separately, must be paid by D.D. payable to municipal

commissioner, Ahmedabad, along with tender document submitted to AMC.Cheque not accepted

Tender No.	Tender floating date	Last date for submission up to 2.00 p.m. and tenders will be opened at 3.00 p.m.
Tender No –T 1 to T-29	20th September 2008	10th October 2008
Tender No –T 30 to T-53		14th October 2008

Municipal commissioner reserves the right to reject any or all tenders without assigning any reason

PUBLISHED BY CENTRAL MEDICAL STORE
OLD PILOT DAIRY BUILDING, KANKARIA
AHMEDAD-380022

N.B. - Tender Conditions & Specifications Must Be Attached Along with The Tender.

AHMEDABAD MUNICIPAL CORPORATION
CENTRAL MEDICAL STORES
Old Pilot Dairy Building, Kankaria,
Ahmedabad. -380022
Telephone No. 079-32981605, 25327571
Phone/ Fax No No.:079-25327571

TENDER FORM

IMPORTANT INSTRUCTIONS / TERMS / CONDITIONS TO BIDDERS FORMING PART & PARCEL OF ENQUIRY DOCUMENT:

PLACE OF SUBMISSION AND OPENING OF THE TENDER
Tender Officer – Record Office
Sardar Patel Bhavan
Mahanagar Seva Sadan, Ahmedabad
Ahmedabad- 380001.

Tender Notice No. & Name	**8, HOSPITAL EQUIPMENTS (T-1 TO T-29)**	
Last Date Of Submission Of Tender	**10th October**	**UPTO 2-00 P.M.**
Date Of Opening Of Tender (Technical Bid)	**10th October**	**AT 3-00 P.M.**
Tender Fee	**Rs.500/-**	**(NON REFUNDABLE)**
Name & Address Of Bidder		

SPECIAL IMPORTANT NOTES: -

This Contract Is Rate Contract And Quantity Basis. (Quantity May be varied)
All Photo Copies Must Be Submitted Duly Certified True Copies.
No Document Will Be Accepted After Opening Of Technical Bid.
Municipal Commissioner, Ahmedabad Reserves The Right To Accept Or Reject Any Or All Tenders In Part Or Full Or Close The Tender Enquiry Without Assigning Any Reason, At Any Time, At Any Stage.
Technical Bid Must Be Submitted In Duplicate And Bidder Has To Mention Original Copy And Duplicate Copy Very Clearly.
The Specifications Available In The Equipment Offered Shall Be Clearly Mentioned Against Each Tender Specification.
For Any Imported Equipment/Part Of Equipment, Valid Documentary Evidence Regarding Foreign Make As Well As Import Of Goods Is To Be Provided By The Tenderer Along With Consignment.
Any Item Material, Material, Consumable Required For Installation And Commissioning Of Equipment To Be Brought By Supplier At The Time Of Installation Of Equipment.

AHMEDABAD MUNICIPAL CORPORATION
CENTRAL MEDICAL STORES
Old Pilot Dairy Building, Kankaria,
Ahmedabad. -380022
Telephone No. 079-32981605, 25327571
Fax No./Phone No.:079-25327571

TENDER FORM

IMPORTANT INSTRUCTIONS / TERMS / CONDITIONS TO BIDDERS FORMING PART & PARCEL OF ENQUIRY DOCUMENT:

(INST) (PART-I)
GENERAL

1 This document comprises of FOUR parts labeled as Part-I/II/III/IV & Schedule as under:

a) **Part I:** Terms / Conditions & Important Instructions to the Bidders regarding tender / supply of various items:
b) **Part II:** Summary of the offer, and also contains the format of certificates (as applicable);
c) **Part III:** Offer documents for Technical Bid: **(In duplicate).**
d) **Part IV:** Offer document for Commercial Bid: **(In duplicate).**

2 Parts I / II/ III & Schedule with the relevant documents / certificates should be submitted in one sealed envelope marked **"Technical Bid"**
Part IV should be submitted in a separate sealed envelope marked **"Commercial Bid"** which should contain offer as per Drug Part-IV.
Demand draft of E.M.D., Demand draft of Tender Fee & Bank details (as per annexure- V) to be submitted with another envelope marked **"E.M.D.".**
All the envelopes should be clearly super scribed with Tender Enquiry number and the relevant part number (I/II/III or E.M.D.) as the case may be. All the three / four envelopes (as the case may be) should be submitted in one envelope on which the tender enquiry #, name of Bidder and date of opening of tender should be clearly written.

3 **Each page & all documents submitted along with Bid must be signed and stamped by authorized signatory.**

4 No tender will be accepted after prescribed closing time for submission of the same. The delay will not be condoned for any reason whatsoever. However, if the last date of submission of tenders is declared as a holiday by the AMC / Government, the last date of submission of tenders will be extended to the next working day with the same time and venue specified in the tender.

5 Bidder should read this document very carefully and comply with the instructions / terms / conditions therein. Any tender which does not confirm with the instructions / terms / conditions therein is liable to be rejected without any clarification.

6 In case, tender quoted by Distributor than authority letter of manufacturer should be submitted. In case when the authority letter is submitted by subsidiary company of their foreign manufacturer, the subsidiary company will have to produce documents that they are 100% subsidiary company of their foreign manufacture whose items / instruments is quoted.

7 In case where item is required with ISI Mark, manufacturing and selling experience of product having ISI mark will only be considered. But, in such a case, certified copies of licenses to use ISI mark for last one year as well as latest and valid license are to be submitted with tender. The license should be valid during the period of rate contract. It will be the sole responsibility of R.C. Holder to ensure that necessary certificate and permission must be valid till the expiry of the Rate Contract.

8 The certificate should be in force and valid on the last date of the tender / signing of the agreement {as the case maybe}. In case, the certificates / licenses / permission are outdated or the validity period is over, the proof of applying for renewal should also be attached. Such certificates will be considered if the renewal has been applied for within the time limit prescribed for the renewal of that permission / license / certificate under the relevant rules and further if such application for renewal is not specifically rejected by the competent authorities. In case any certificate is still awaited from the competent authority, the proof of making the application should also be attached which will be considered if the application is not specifically rejected by the competent authorities.

9 Variation in the terms and conditions of the tender, including deviation from standards / specifications / terms of supply, payment will not be accepted. **No conditional offer / quotation will be accepted.**

10 In Part III / IV no modification should be done by the Bidder in the name of item, Code No and in the specifications / grade / quality of the item given in the enquiry document. **In Part III (Technical bid)** all information asked / required including specification available in offered models should be mentioned clearly. Additional specifications / features if any available in the offered models shall be provided in additional sheet. **Part IV the rates, applicable taxes (GST / CST / VAT), and total rates with applicable GST/CST/VAT of the item should be filled in along with the item.** In both the cases, nothing else should be written or filled in either Part III/IV.

11 The rates should be distinctly stated. The tax viz. Sales Tax. Excise, etc. should be clearly mentioned, if nothing is stated the rate will be treated inclusive of all taxes and treated as final rate, but bidder has to give split of rate if demanded from Central Medical stores, Ahmedabad.

12 In no case, rates should be quoted anywhere except in part IV. The tender will be summarily rejected without any further processing or reference if the rates are quoted or written at any place except at the relevant place in Part IV.

13 In no case, rates for any item should be quoted in Part IV for which the technical details are not given or the items not quoted with details. In that case, rate quoted for that item will not be considered.

14 **E.M.D. & TENDER FEE:**

Individual Demand draft to be paid for E.M.D. & Tender Fee in favor of **"Municipal Commissioner, Ahmedabad".** Bidder has to submit separate D.D. as tender fee tender vise. Tender Fee Rs. 500/=per tender is non refundable.

Any Tender without E.M.D. and Tender fee would be outright rejected.

If the tender is accepted, bidder has to submit security deposit at the time of signing the rate contract. The amount of security deposit will be 5% of the contract value (in rounding of thousand value) and it will be retained till the due performance of the contract. After the due period of the contract the amount of security deposit (S.D.) will be refunded on production of no objection certificate from the direct indenting officer.

Bidder will have to inform to Central Medical stores, Ahmedabad if their Bank details changed after submission of bid. Other wise in case of E.M.D. or Security Deposit return, payment would be done on the basis of bank details they have submitted with bid.

The E.M.D./Security Deposit shall liable to be forfeited in the following circumstances when the:

(a) Bidder fails to sign the agreement for entering into contract in case the offer is accepted, due to any reason whatsoever.

(b) Bidder fails to replace the goods with in given stipulated time, declared **not of standard quality** or not conforming to acceptable standards or found to be decayed / infected / spoiled before the date of expiry.

15 Bidder will have to submit ten-valued customer's list & Product literature of all quoted products.

16 (i) The detailed required specifications of the equipment to be purchased are shown in technical Bid. The tenderer should either confirm whether the items offered are conforming to these specifications clearly OR specifications available in offered product should shown against each specification along with name of manufacturers. Brand / Model only in the format provided in technical bid. Please note that tenderer should not write 'As per Literature enclosed '. In such a case, offer shall not be considered.

(ii) The Tenderer should have to arranged demonstration of equipment offered as and when desired at their expense failing which tender shall liable to be ignored. In case where samples are required to be submitted, it should be kept ready and submitted as and when demanded in duly sealed pack super scribed with details of tender enquiry No, due Date, Name and address of tenderer, make / model etc. failing which offer shall liable to be rejected.

17 The rates quoted in part IV must be F.O.R. destination at C.M.S. Ahmedabad and Ahmedabad Municipal corporation's Hospitals with installation per unit basis mentioned in the enquiry document and should be comprehensive incorporating the cost of the instrument / equipment and accessories required as part of the equipment and shown as such in the enquiry document. In case any item is required as an essential accessory for operating the equipment, it must be mentioned clearly in Part-III and its rates must be included in the rates for the equipment in Part-IV. If not mention of such essential accessories is made in Part-III or if its rates are not shown in Part-IV, it will be presumed that the cost of essential accessories is included in Part-IV and no separate payment for the same will be made thereafter under any circumstances. If any accessory is demanded as mandatory in the tender enquiry, under no circumstances it should be shown as an optional accessory and quoted separately. The price of the product offered must include the accessories required for operation of the instrument and no separate payment will be made even if such an accessory is not included in the quotation. Only those accessories which are specifically recommended by the manufacturer of the instrument / equipment should be offered as part of the equipment and under no circumstances a cheaper variety of an accessory not approved / recommended by the manufacturer should be offered.

18 Tenders will be opened on the date, time & venue specified in tender notice and the Bidder or his authorized agent has the liberty to be present at the time. However, in unavoidable circumstances, the Municipal commissioner or Demanding officer appointed by him can change the date / time / venue of opening or open the tenders in the absence of Bidders or his agent at his discretion. In case the Government/A.M.C. declares the specified date as holiday, the tenders will be opened on the next working day at the same time and venue specified in the tender.

19 (a) In the first instance only the **"Technical Bid"** envelope containing Parts I/II/III of the documents will be opened on the date of opening of the tender and taken into consideration for finalization. Subsequently, the **"Commercial Bid"** envelope containing Part IV will be opened only of those Bidders who's Technical Bid satisfy the technical requirements of the tender and are otherwise acceptable. No intimation for commercial bid would be given to Bidder whose technical bid disqualified.

(b) Technical bid and commercial bid to be submitted original copy and duplicate copy and in case of any discrepancy in original copy and duplicate copy the original copy will be final and binding.

(c) Once the tender is submitted it will be the responsibility of the Bidder not to escape half way directly or indirectly by way of raising any problems

20 Municipal Commissioner reserves the right to accept similar tender quoted item equivalent to specification or other than the specified one, may be accepted.

21 **TERMS OF SUPPLY**

RISK PURCHASE: If the contractor fail to deliver the stock or any part thereof within the period prescribed for delivery to the Municipal Commissioner or the Direct Demanding officer or any officer authorized by Municipal Commissioner in his behalf, may take any of or more of the following steps:

(a) Recover from the contractor as liquidated damages a sum equivalent to one half percent of the price of the undelivered stock at the stipulated rate for each week or part there of during which the delivery of stock delayed, subject to a maximum limit of 10% in case of order less than Rs. 1 lac and 5% in case of order exceeding Rs. 1 lac value.

(b) To purchase undelivered stock or others of a similar description where others are exactly complying with the particulars in the opinion of Municipal Commissioner or Demanding officer or any officer authorized by him on his behalf (which shall be final) on contractors

account and at contractor's risk.

(c) To cancel the contract on the event of action being taken under a) and b) above and contractor shall be liable to make payment for any loss Municipal corporation sustain by reason of higher price of the stocks so purchased.

22 On the event of order being placed with the contractor against this tender, if contractor fails to supply any stocks as per terms and conditions of acceptance of tender or fails to replace any stock rejected by the direct demanding officer or Any person authorized by Municipal Commissioner on his behalf they are entitled to purchase from elsewhere such stocks from any other source at such price at his discretion without notice to the contractor and if such price shall exceed the rate set out in the schedule of acceptance of tender, the contractor shall be responsible to pay the difference between the price at which such stocks have been purchased by the direct demanding officer and the price calculated at the rate setout in the schedule.

23 In the event of breakages or loss of stocks during transit against requisition order the said quantity will have to be replaced by the R.C. Holder.

24 **Delivery Period:** Unless specified in the tender or instructed by municipal commissioner, Ahmedabad or Direct demanding officer maximum delivery period shall be of four weeks.

25 The payment of the bills shall be withheld in the following circumstances:

26 The goods are found sub-standard or in non-acceptable conditions:

b) Breach of any condition of contract by the R.C. Holder

27 The claim of price revision of finished goods will not be allowed under any reason, in ordinary circumstances except government taxes, excise duty, justification with Govt.order.

28 The installation of equipment is to also to be completed within 30 days of delivery.

29 Tenderers will have to provide four free services for maintenance during the warranty period. In addition to provide services, supplier has to attend the complaint, if any, for any defects, within 48 hours including replacement of any defective part, failing which Rs.100/day (in case of item value is less than 5 lac) and Rs.300/day (in case of item value is more than 5 lac) will be recovered and period of break down of equipment shall be excluded from the warranty period. In such a case period of warranty shall be extended accordingly. After completion of warranty period, if the Municipal commissioner, Ahmedabad or Direct demanding officer on his behalf desires, supplier will have to enter in to a comprehensive maintenance contract for 5 years at the rate per annum not exceeding 10 % of the total contract would include free replacement of any part / component, if required, during the period of the contract. The maintenance contract of the above shall be made operative from any time during the first five years from the date of contract. Any offer for service contract contrary to these conditions will not be accepted. Tenderers will also have to guarantee regular and timely supply of all the spare parts required for the normal working of the item at a cost not exceeding the cost of such a part charged to any other Government organization.

30 Suppliers are required to impart training to operate equipment at the place of installation within seven days of its installation.

31 **Replacement Clause**: If any stocks supplied to Municipal General Hospitals or Central Medical Stores, Ahmedabad against the R.C. are found to be "**not of standard quality"** on inspection and / or analysis the entire quantity has to be replaced within three weeks or make full payment of the entire consignment (Batch) against the particular invoice irrespective of the fact that part or full quantity (Batch) of the Stocks Supplied have been consumed**.**

32 Bidders convicted of an offence under the Drugs Act. 1940, the Dangerous Drugs Act. 1930 and the Bombay Prohibition Act, 1949 or the Bombay Opium Act, 1938 or any related act in force from time to time as adopted by Gujarat State will not be eligible for being given contracts.

33 A Bidder must enjoy confidence of the Commissioner Drugs Control Administration, Government of Gujarat and the Drugs Controller, Government of India and must not have been prosecuted for Sub standard products.

34 Bidders should see that the words " A.M.C." are either printed or rubber stamped on the labels pasted on each container by the manufacturing firms or the suppliers.

35 Bidders should submit the tender duly sealed (by sealing wax) and super scribed properly, failing which the tender will be rejected.

36 Validity: Bidders are requested to give 180-calendar days validity period for their offers from the date of opening of the technical bid.

37 Bidder whose tender is accepted will have to execute an agreement on stamp paper at his cost.

38 In case of a partnership firm and a joint Stock company the contract must be signed by the Directors or by a person duly authorized to sign the contract for the company or firm.

39 Bidder shall have to supply goods to different medical institutions in the Municipal Corporation limits. The invoice of supplies to such institutions must be sent in four copies, of which three copies to the indenting officer and one to the Asst.Manager, Central Medical Stores, Municipal corporation, Ahmedabad.

40 Any kind of change in the dept. like dismantling & refixing of the fall ceiling, fixing of plate on ceiling/strong part of any horizontal beam, etc. Must be done by the tenderer, without disturbing the look of the dept.

41 The payment against supply of stocks shall be made within 30 to 45 days from the date of receipt of the goods in good and acceptable conditions at the destination (in case of surgical instrument after successful installation). However, no interest will be chargeable by the R.C. Holder, if the payment is delayed.

42 Direct Demanding Officers: All Superintendents of the general hospital of Municipal corporation and Asst. Manager, Central Medical Stores, Ahmedabad declared as Direct demanding officers.

43 Warranty & guarantee period should be mentioned clearly.

44 The issue of this contract is a rate contract and quantity basis. Each and every order received during the currency of the contract from the direct demanding officer should be executed by the contractor, who shall held an adequate stock for the purpose.

45 The rate contract can be terminated by the Municipal commissioner, Ahmedabad in the following circumstances:

a) If the firm is debarred or disqualified or ceases to exist or convicted of any offense.

b) If the quality of the item to be supplied is found not up to the **standard quality** and multiple samples are found to be **"not of standard quality"**.

c) If the supply position of the firm is not satisfactory.

d) On breach of contract.

46 Bidders should quote product of **one company only** in each column, failing which tender may be rejected.

47 The period of the rate contract shall be one year from the date of issuance of Rate Contract. However, the Municipal Commissioner reserves the right to terminate the contract at any time at any stage without assigning any reason by giving one-month notice.

48 For any clarification Bidders may contact Asst. Manager/O.M.S., Central Medical Stores during 3 p.m. – 4 p.m. on any working day at Central Medical Stores, Ahmedabad.

49 This contract is subject to Ahmedabad jurisdiction only.

ACCEPTANCE OF TENDER:

The Bid will be rejected due to any reason or reasons mentioned below:

1. Non-submission of tender within stipulated time.
2. Submission of tender without tender fee and E.M.D.
3. Bid unsigned or unstamped or not initialed on each page or with unauthenticated correction**.**
4. Submission of tender documents in unsealed envelope.
5. Bid not submitted in separate envelopes as per conditions and the envelopes are not super scribed with details of the tender enquiry and part enclosed.
6. Non-submission of required documents.
7. Conditional and / or vague offers.
8. Unsatisfactory past performance of the Bidder.
9. Rates have been shown elsewhere than **Part IV.**
10. Items with changes / deviations in the specifications / standard / grade / packing / quality.
11. Submission of misleading / contradictory / false statement or information and fabricated / invalid documents.
12. Tender not filled up properly/hand written tender.
13. Offering a cheaper accessory not approved / recommended by the manufacturer.
14. Non-submission of authority letter in prescribed format in case of suppliers/authorizes distributor.
15. The legible and certified copies of the following documents **NOT** attached / annexed to **Part II:**
    a) Valid Manufacturing license & Product Permission (as the case may be), for imported products valid Manufacturing License & Product Permission of manufacture (as the case may be);
    b) Valid Appropriate Drug License of manufacturer & distributor & Valid Import License for imported Products.
    c) Valid (Good manufacturing practice) G.M.P. / W.H.O. G.M.P. / D.G.Q.A. certificate, for imported drug products valid G.M.P. / W.H.O. G.M.P. certificate of manufacturer;
    d) I.S.O. certificate (If applicable).
    e) Valid I.S.I. certificate (if applicable).
    f) Non-conviction certificate (Recent).

(INST) (PART-II)
ANNEXURE-I
Verification, Undertaking & Documents

From: M / s. ---------- No.----------

To: Asst.Manager/O.M.S.
Central Medical Stores
Old Pilot Dairy, Kankaria,
Ahmedabad.

Sub: Supply of Drugs / Medicine / Surgical goods etc.

Ref: Tender Enquiry // ----------

Sir,

I / we enclose the priced tender from duly signed & stamped. Necessary documents are enclosed herewith. The rates quoted against each item are inclusive of excise duties and other charges. I/We shall abide by all these conditions. I/We hereby undertake to supply the goods to the consignees as per the terms of the R.C Mentioned in the Bid documents.

I/We certify that the copies of the certificates / documents attached herewith are authentic true copies of the original certificates / documents and I/We undertake to produce the original certificates / documents for verification on demand. I/We undertake to supply the attested copies of certificate / documents required at the time of signing the letter of agreement if my/our offer is accepted.

I/We certify that I/We are in possession of the requisite licenses / permits required for the manufacturer / supply / sale / distribution of the items and further certify that the said licenses / permits have not been revoked / cancelled by the issuing authorities and are valid as on date. I/We also certify that I/We have not been blacklisted / debarred / stopped form supply of the items quoted in the Tender Enquiry by any Government Department / Organization / undertaking in India.

I/We also take cognizance of the fact that providing misleading or questionable information or failure to furnish correct or true information to you or any other Direct Demanding Officer or failure to comply with any contractual requirement laid down by you / them will be considered as a serious breach of the terms and conditions of the tender and will invite disqualification and other penal action as deemed fit by the Authority.

Yours faithfully,

Date: ----------

{Signature & Stamp of Bidder}

(INST) (PART-II)

ANNEXURE-II

(FORMAT OF MANUFACTURING EXPERIENCE & MARKETING DATA)

I _____ age_____ residing at _____ in capacity of _____ M/s. _____ _____ here by solemnly affirm that …

We have manufacturing experience of tender quoted Item No. ---------- for _____ year's

Sr. No.	Item code	Name of Item	YEAR	Continuance of production & marketing of last three years		
				Mfg./ Imported Qty.	Sold Qty.	Sale Value (Rs. Lac)
			2005-06			
			2006-07			
			2007-08			
			2005-06			
			2006-07			
			2007-08			
			2005-06			
			2006-07			
			2007-08			
			2005-06			
			2006-07			
			2007-08			

2. That on the basis of the above facts & figures we M/s. _____ has manufactured / marketed Our products prior to the due date of Tender Enquiry No._____ as per the specifications mentioned in the tender.

Whatever stated above is true & correct.

(If required attach photocopy of sheet)

{SIGNATURE AND STAMP OF AUTHORISED SIGNATORY}

Date: _____

(INST) (PART-II)
ANNEXURE-III

(FORMAT OF AUTHORIZATION LETTER FOR AUTHORIZED DISTRIBUTOR) (IN ORIGINAL)

I / we _____
hereby declare that…

1 M/s _____ is our authorized distributor for our products from date _____ and they are authorized to quote and follow up on our behalf and the said agreement is valid in force as on date;

2. I / We undertake to supply the drugs / items for which the quotations of following items are submitted by M/s, _____ on our behalf in respect of Tender Enquiry # _____

Sr. No.	**Item No.**	**Name of Item**
(1)		
(2)		

3. I / We have read all the terms and conditions of the enquiry and the same are irrevocably binding upon us till the expiry of the contract signed & executed on our behalf;
4. I / We shall notify the Municipal commissioner, Ahmedabad immediately if there is any change in the agreement between M/s. _____and me/us regarding authorized distributorship of our products and further undertake to supply the items quoted by the distributor on my / our behalf at the quoted in the tender enquiry in case of such a change of agreement.
5. This authority is applicable only for Tender Enquiry #_____

Date:

{Signature & Stamp of Bidder}

(INST) (PART-II)
BANK ACCOUNT DETAILS
ANNEXURE –IV

1	**Name of the party**	
2	**Address**	
3	**Telephone No./ Mobile No.**	
4	**Email ID**	
5	**Pan No.**	
6	**Name of the Bank**	
7	**Branch & Address of Bank**	
8	**Account Type**	
9	**Bank Account No.**	
10	**MICR Number**	
11	**I.F.C.E. Code No. (Bank)**	
12	**Amount Payable**	

My above bank details are true and correct.

Bank details shall be providing on bidder's letter pad and to be submit in E.M.D.Envelope **only.**

{Signature & Stamp of Bidder}

(INST) (PART-II)
CHECK LIST

		Yes/No	N.A.	Your's Page No.
1	D.D. of tender fee submitted			
2	D.D. of E.M.D.submitted.			
3	Bid stamped and initialed on each page**.**			
4	Submission of tender documents in sealed envelope.			
5	Separate envelope of EMD.			
6	Separate envelope of Technical bid.			
7	Separate envelope of commercial bid.			
8	Authority letter			
9	Valid Manufacturing license & Product Permission (as the case may be),			
10	For imported products valid Manufacturing License & Product Permission of manufacture (as the case may be);			
11	Valid Appropriate Drug License of manufacture & distributor & Valid Import License for imported Products.			
12	CE certificate			
13	I.S.O. certificate (If applicable).			
14	Valid I.S.I. certificate (if applicable).			
15	Non-conviction certificate (Recent).			
16	Each envelopes super scribed with details of tender inquiry and part enclosed.			

(INST) (PART-III)
PART III: TECHNICAL BID
(IN DUPLICATE)

Name of Item**: Defibrillator**		**Tender No: T 1**
Manufacturer:	Brand:	Model:
Bidder Name:		**E.M.D.:Rs.25000/-**
Required Specifications:		Specifications available in offered model
➢ Dual mode of power as well as Battery operated		
➢ Energy: 0 to 360 Jules with display on screen.		
➢ Charge Time: <5 seconds with battery and <15 seconds on AC.		
➢ Charge Indicator: Visual and audible tones		
➢ Synchronization: Synchronization to R wave with marker indication on ECG waveform.		
➢ Paddles: Standard adult anterior electrodes slide off to expose paediatric with 600 mm coiled cable electrodes.		
➢ Monitor: LCD with built in cautry filter and on screen message tone for leads off.		
➢ Printer: Thermal with annotations like Hospital name, Time, Date, Heart- rate, ECG Parameters, Defib. Mode selected and delivered energy, patient impedance, Peak current and Hospital name.		
Standard Accessories:		
o Paddles adult & paediatric		

o Disposable electrodes	
o Patients cable [3 leads]	
o Mains cable [Power cord]	
o Internal paddles	
o Cardiac jelly	
o Earth cable	
o Accessories bag	
o User manual	
o Recording paper	
o Trolley.	

(INST) (PART-III)
PART III: TECHNICAL BID
(IN DUPLICATE)

Name of Item: **Syringe Infusion Pump**		**Tender No: T 2**
Manufacturer:	Brand:	Model:
Bidder Name:		**E.M.D.:Rs 3000**
Required Specifications		Specifications available in offered model
➢ With single and dual syringe attachment.		
➢ Battery as well as mains operated.		
➢ Facility to deliver adjusted rate & volume and visual display with safety system.		
➢ Transparent porfuror tubing portable.		

(INST) (PART-III)
PART III: TECHNICAL BID
(IN DUPLICATE)

Name of Item: **Pulse Oxymeter with ECG Monitor**		**Tender No: T 3**
Manufacturer:	Brand:	Model:
Bidder Name:		**E.M.D.:Rs.3000/-**
Required Specifications:		Specifications available in offered model
➢ 5" to 7" LCD with backlit display		
➢ Separate visual alarm indicator for ECG and SPO2		
➢ Built in rechargeable battery backup		
➢ Low and high alarm units		
➢ Real time display		
➢ Twenty four hours storage of data		
➢ SPO2 probe for adult as well as paediatric.		

{Signature & Stamp of Bidder}

INST) (PART-III)
PART III: TECHNICAL BID
(IN DUPLICATE)

Name of Item: **Pulse Oxymeter**	**Tender No: T 4**
Manufacturer: Brand:	Model:
Bidder Name:	**E.M.D.: Rs.1000/-**
Required Specifications:	Specifications available in offered model
➢ 5" to 7" LCD with backlit display	
➢ Trend of SPO2 and pulse rate for last 24 hours.	
➢ Plethysmograph display	
➢ Visual and audible alarm for low and high value	
➢ Built in rechargeable battery.	

(INST) (PART-III)
PART III: TECHNICAL BID
(IN DUPLICATE)

Name of Item: **Five Function Multipara Monitor**		**Tender No: T 5**
Manufacturer:	Brand:	Model:
Bidder Name:		**E.M.D.: Rs.3000/-**
Required Specifications:		Specifications available in offered model
➢ Five function monitor		
➢ 10" TO 12" color display		
➢ Arrhythmia detection		
➢ Pace maker detection function		
➢ Core & Peripheral temp monitoring		
➢ Minimum 24 hours trend storage function		
➢ Battery back-up		
➢ Visual and audible alarms for all separate function		
➢ BP cuff for adult, large adult, paediatric and neonate [Reusable]		
➢ SPO2 probe for adult and paediatric.		

(INST) (PART-III)
PART III: TECHNICAL BID
(IN DUPLICATE)

Name of Item: **Capnograph With Pulse Oxymeter**	**Tender No: T 6**
Manufacturer: Brand:	Model:
Bidder Name:	**E.M.D.:Rs.5000/-**
Required Specifications:	Specifications available in offered model
5-6" LCD display with backlight	
Battery operated	
Battery life should be more than 2 hours.	
Capnograph and plethysmograph	
Pulse oxymeter probe for adult and child	
Should show ETCO2 respiration rate, pulse rate and saturation of O2. Response time should be 1.5 sec. Trend length 4-8 hours.	
Side stream OR recent technology of micro stream.	
Light weight and portable	
Visual and audible alarms	

(INST) (PART-III)

PART III: TECHNICAL BID
(IN DUPLICATE)

Name of Item: **Multipara Monitor**		**Tender No: T 7**
Manufacturer:	Brand:	Model:
Bidder Name:		**E.M.D.:Rs.5000/-**
Required Specifications:		Specifications available in offered model
➢ 5-6" color TFT display		
➢ ECG, SPO2 and NIBP		
➢ Adult as well as paediatric probe for SPO2		
➢ NIBP cuff for adult, large adult and paediatric [Reusable]		
➢ NIBP cuff for neonate disposable		
➢ Battery back-up		
➢ Data storage for 24 hours		
➢ Graphic and tabular trend information, systolic, diastolic and ➢ mean ready for NIBP		
➢ Light weight and portable		

➢ Visual and audible alarms	
➢ ECG Cable with 5 leads.	

(INST) (PART-III)
PART III: TECHNICAL BID
(IN DUPLICATE)

Name of Item: **Vital Sign Monitor**		**Tender No: T 8**
Manufacturer:	Brand:	Model:
Bidder Name:		**E.M.D.: Rs.5000/-**
Required Specifications		Specifications available in offered model
Multi parameter patient monitor with facilities for measuring following parameters:		
➢ ECG [5 -12 lead mode]		
➢ Respiration		
➢ SpO2		
➢ NIBP		
➢ Temp		
➢ Heart Rate Monitor		

(INST) (PART-III)
PART III: TECHNICAL BID
(IN DUPLICATE)

Name of Item: **Ventilator (Adult)**		**Tender No: T 9**
Manufacturer:	Brand:	Model:
Bidder Name:		**E.M.D.: Rs.3000/-**
Required Specifications		Specifications available in offered model
➢ With OR without dual display (Graphic, machine, patient)		
➢ Tidal Volume: 50-1500 ml.		
➢ Peak flow: 180 ltr./ min.		
➢ Frequency: 8-40 BPM		
➢ SIMV Frequency: 1-40 BPM		
➢ PEEP: 0.20 Mb		
➢ Pressure support: 1-30 Mb		
➢ FIO2: 21-100%		
➢ Inspiratory pressure trigger: - 1 to 20 Mb		
➢ I/E Ratio: 25-50%		
TECHNICAL:		
➢ Electric Operated		
➢ Medical gases via central connection / bulk cylinders.		
➢ Air compressor driven		
MODE:		
➢ CMV		

➢ ACMV	
➢ APCV	
➢ PSV	
➢ CPAP	
➢ NIV	
➢ SIMV in volume control & pressure control both	
➢ PCV	
ALARM:	
➢ Pressure alarm	
➢ 02 alarm disconnection	
➢ Apnea alarm	
➢ Lower supply failure alarm	
➢ Battery back up for power failure	

(INST) (PART-III)

PART III: TECHNICAL BID
(IN DUPLICATE)

Name of Item: **Computerized Stress Test System (T.M.T. Machine)**	**Tender No: T 10**
Manufacturer: Brand: Model:	
Bidder Name:	**E.M.D.: Rs.15000/-**
Required Specifications:	Specifications available in offered model
➢ System should acquire and analyze 12/15 leads.	
➢ System should have wireless transmission of patients ECG to analyzer, preferably telemetry.	
➢ Should provide standard full Interpretation of Supine ECG with reasoning.	
➢ Display of real time 12 lead diagnostic qualities ECG waveform, average complexes beat of all 2 leads with color comparison along with digital value of ST level and slope.	
➢ Display of four times enlarged complex with color-coded comparison with reference and facility for dynamic lead selection for maximum ST changes.	
➢ System should display running trend of ST, Heart Rate and Mets on the screen to evaluate patient condition during exercise phase.	
➢ System should have an ability to readjust J-ST interval measurement points and generate re-analyzed new reports of multiple patients data stored in the system.	

➢ It should display graph on the monitor similar to the graph on the recording paper to help easy measurement of ST changes during the test.	
➢ Should have full disclosure playback review and storage for minimum of 90 tests.	
➢ System should provide Multiple and customizable printing formats as per user's preference. Facility to export report in PDF format for review on any normal computer for emailing.	
➢ System should have muscle artifact filter 40 Hz low pass, 50 Hz line filter. It must have additional filter, which reduces noise without affecting the ECG frequency response.	
➢ Reports should be printed in plain paper on Laser Printer and / thermal printer.	
➢ Heavy Duty Treadmill: Heavy Duty noise free TREADMILL with speed ranging from 0.5 to 12 mph and grade of 0-25%. Imported Heavy Duty taskmaster.	
➢ Compatible programmable Non Invasive blood pressure measurement module with Bi-directional transfer of data between module and analyzer.	
➢ Equipment should comply with FDA & CE rules and regulations.	
➢ Company quoting the tender must be ISO certified.	

(INST) (PART-III)
PART III: TECHNICAL BID
(IN DUPLICATE)

Name of Item: **Twin dome Ceiling Suspended surgical O.T. Light** **Tender No: T 11**	
Manufacturer: Brand: Model:	
Bidder Name: **E.M.D.:Rs.25000/-**	
Required Specifications:	Specifications available in offered model
➢ Should be of international standard.	
➢ Valid safety Certification from competent Authority.	
➢ Should possess 601 E certification	
➢ Should have CE certification	
➢ Input Supply 230 Voltage	
1. One Major dome	
➢ Intensity of major dome light should not be less than 125000 luxs	
➢ Diameter of major dome 760 mm if seven bulb, if one bulb then 700 mm	
➢ Bulbs 1 or 7 halogen	
➢ With Filler	
➢ Color Temperature 4200 -4300 k	
➢ Focusing	

➢ Field size -200 mm diameters.	
2. One minor dome	
➢ Intensity of minor dome light should not be less than 750000 luxs	
➢ Diameter of minor dome 630 mm if five bulbs. If one bulb 500 mm.	
➢ Bulbs 1 or 5/4 halogen bulb	
➢ with Filter	
➢ Color Temperature 4200-4300 k	
➢ Focusing manual	
➢ Field size minimum 200 mm diameters.	
➢ Rotation at ceiling end 360 degree.	
➢ Rotation at spring arm. 320 degree.	

(INST) (PART-III)
PART III: TECHNICAL BID
(IN DUPLICATE)

Name of Item: **Sonic Ultrasound Scalpel**		**Tender No: T 12**
Manufacturer:	Brand:	Model:
Bidder Name:		**E.M.D.:Rs.15000/-**
Required Specifications:		Specifications available in offered model
Hardware		
Generator		
Foot switch		
Cart		
Standard Accessories consisting of:		
Hand piece		
Blade system adaptor, 5 mm		
Shears Adaptor		
Hand switching adaptor		

{Signature & Stamp of Bidder}

Blade wrench	
➢ Open Surgery Instruments:	
Coagulating shears, 10 mm	
Short curved coagulating shears, 5 mm	
Short curved coagulating shears, 8.5 mm, 18 cm long	
Dissecting hook, 5 mm	
Sharp hook blade, 5 mm	
Endoscopic Surgery Instruments	
Dissecting hook, 5 mm	
Curved blade, 5 mm	
Laparoscopic Coagulating shears, 10 mm	
Laparoscopic Coagulating shears, 05 mm, knife mode	
Laparoscopic Coagulating shears, 05 mm dia., curved mode, 36 cm long	
Laparoscopic Coagulating shears, 05 mm dia., curved mode, 45 cm long	
Laparoscopic Hand activated Coagulating shears with Clicker diameter, Curved mode, 36 cm long	
Training should be provided to surgeons/doctors from hospital.	

(INST) (PART-III)
PART III: TECHNICAL BID
(IN DUPLICATE)

Name of Item: **Radio Lucent Operation Table**		**Tender No: T 13**
Manufacturer:	Brand:	Model:
Bidder Name:		**E.M.D.:Rs.5000/-**
Required Specifications:		Specifications available in offered model
Operation Table for Leparoscopic Procedure		
Automatic Remote controlled		
Single pillar eccentric design		
Height adjustment facility		
Trendelenberg tilt and lateral tilt movement automatic		
With Urology Board and Lithotomy stirrups & std. accessories.		

{Signature & Stamp of Bidder}

(INST) (PART-III)
PART III: TECHNICAL BID
(IN DUPLICATE)

Name of Item: **'0' Degree Adult Size Telescope for Cystoscope**		**Tender No: T 14**
Manufacturer:	Brand:	Model:
Bidder Name:		**E.M.D.: Rs.3000/-**
Required Specifications:		Specifications available in offered model
➢ Adult size		
➢ Imported or Indian make,		
➢ Diameter- 4 mm for lence		
➢ 5.1 mm for outer sheath		
➢ Fiber optic light transmission incorporated		
➢ Straight forward telescope 0 enlarge view		
➢ Autoclavable		
➢ Should fit to all universal standard resectoscope & cystoscope		
➢ Color code green		

(INST) (PART-III)
PART III: TECHNICAL BID
(IN DUPLICATE)

Name of Item: **Leproscopic Three Chip Camera**		**Tender No: T 15**
Manufacturer:	Brand:	Model:
Bidder Name:		**E.M.D.:Rs.5000/-**
Required Specifications:		Specifications available in offered model
Leproscopic Three Chip Camera with all standard accessories including:		
Digital Video Enhancement		
Automatic white balance device for video printer/recorder		
Zoom facility		
On screen display		
Anti moiré filter		
Digital Gamma processor		

(INST) (PART-III)
PART III: TECHNICAL BID
(IN DUPLICATE)

Name of Item: **Digital Video Colposcope**		**Tender No: T 16**
Manufacturer:	Brand:	Model:
Bidder Name:		**E.M.D.:Rs.5000/-**
Required Specifications:		Specifications available in offered model
➢ Micro lens digital color CCD		
➢ Mirror image effect for better diagnosis		
➢ E-Flip = image can be rotated upside down		
➢ High resolution more than 450 CCD lines		
➢ More than 650000 pixels		
➢ Zoom in & Zoom out facility available		
➢ High MCD super bright shadow less light		
➢ Automatic focus [20 cm - 50 cm] & manual focus		
➢ Magnification more than 20x		
➢ Magnification Indicator		

➢ Acetic acid test timer	
➢ Built-in electronic green filter	
➢ Vertex floor level indicator	
➢ BNC/SVHS with digital video out	
➢ Color temperature more than 6000K	
➢ Image management software with image grabber card.	

(INST) (PART-III)
PART III: TECHNICAL BID
(IN DUPLICATE)

Name of Item: **Thermachoice Uterine Balloon Therapy system**		**Tender No: T 17**
Manufacturer:	Brand:	Model:
Bidder Name:		**E.M.D.:Rs.10000/-**
Required Specifications:		Specifications available in offered model
The uterine Balloon therapy system for the treatment of hyper topic endometrial layer. This system is for the treatment of DUB & it helps to avoid hysterectomy- A Major Surgery.		
System specification:		
a. Facility of Display window to show Pressure, Temperature, Time & Instructions.		
b. Therapy steps chart with controller unit for ready reference [No need to refer the manual]		
c. Power requirements: 220-240v; 50/60 Hz; 1A; 3 wire grounded system.		
Main fuses: 20-240 VAC; 5X 20mm TI.6A 250V Sow Blow.		
d. Dimensions: 10.2cm (h) X 41.2cm (w) X 37.0 cm (d)		
Heater Fuses: T5A 250V.		
e. A.U.S.F.D.A. approved silicon balloon catheter with unique contouring capability that can fit different size & type of Uterus.		
f. Circulatory propeller to circulate heated fluid through out the cavity uniformly		

g.	There should be provision of Safety in system. The device should automatically deactivate when temperature & pressure parameters go beyond recommended range.	
	Pressure 90-200 mmhg & Temperature 70 to 90C	
h.	There should be alarm facility. The machine should have in built mechanism to give alarm related to connections, change in step, defective catheter/ Umbilical cord, Over/ Under Heating, Over/ Under Pressure.	
i.	The system should have U.S.F.D.A approval.	

(INST) (PART-III)
PART III: TECHNICAL BID
(IN DUPLICATE)

Name of Item: **Leproscopic Instruments** **Tender No: T 18**	
Manufacturer: Brand: Model:	
Bidder Name: **E.M.D.: Rs.10000/-**	
Required Specifications:	Specifications available in offered model
Single chip digital camera, Imported	
Leproscope, 0 degree telescope, Imported	
Leproscope, 30 degree telescope, Imported	
➢ CO2 Insufflator with pre-heating of gas, 16-20-lit/min. flow rate, Imported.	
Xenon Light source 180 W., Imported/Indian.	
Fibre optic cable 4.3 mm, 230 cm., Autoclavable, Imported	
LCD Monitor, 14" screen, Std. Co. make	
Video Recorder, Indian	
HAND INSTRUMENTS, Indian	
1) Veress Needle, 2 mm : 2 pc	

2) Sheath/Cannula flap valve with stopcock, 11 mm	: 4 pc	
3) Sheath/Cannula flap valve with stopcock, 5.5 mm	: 4 pc	
4) Reducer from 11 mm to 5.5 mm	: 2 pc	
5) Trocar with pyramidal, sharp tip, 10 mm	: 2 pc	
6) Trocar with pyramidal, sharp tip, 05 mm	: 2 pc	
7) Traumatic maxi grip grasping forceps, 5 mm	: 2 pc	
8) Dissecting & Grasping forceps, 5 mm	: 6 pc	
9) Atraumatic Fenestrated Grasper, 5 mm	: 2 pc	
10) Metzenbaum curved scissor, 5 mm	: 2 pc	
11) Unipolar connection cable	: 2 pc	
12) Suction Irrigation Cannula, 5 mm	: 2 pc	
13) Clip Applying forcep, Ratable, Detachable.	: 2 pc	
14) Hook electrode, 5 mm, L-shape	: 2 pc	
15) Mayoma screw, 5mm	: 2 pc	
16) Claw forcep, 10 mm	: 2 pc	
17) Laparoscopic Needle Holder, Str/curved	: 2 pc	
18) Fan Retractor, ratable	: 2 pc	
19) Cholangiography forcep, 5 mm	: 1 pc	
20) Cholangiography catheter, 5 mm	: 1 pc	

(INST) (PART-III)
PART III: TECHNICAL BID
(IN DUPLICATE)

Name of Item: **X-Ray Machine 500 mA**		**Tender No: T 19**
Manufacturer:	Brand:	Model:
Bidder Name:		**E.M.D.:Rs.20000/-**
Required Specifications:		Specifications available in offered model
[A] X-Ray Generator		
➢ Radiographic rating 500 mA and 125 kVp.		
➢ Large focus: 100 mA, 150,200,300,400,500mA (multiple station)		
➢ Small focus: 100 mA, 150,200,300,400,500mA (multiple station)		
➢ Radiographic kVp: 40 -125 kVp, in steps of (2-5) kV variation		
➢ Exposure time: Solid state electronics timer from 0.02 to 5.0 sec.		
➢ Fluoroscopic rating: mA, 0.5mA steps, kV 40-100 kV.		
➢ Tube head: 20/40 rotating anode x-ray tube		
➢ HT cable: 1 pair of 6 meter HT cable.		
➢ Digital display: kV, mA, mAs		

➢ Indicator: Main on, overload, and x-ray on by neon lamp	
➢ LV Compensation: +/- 20% of input voltage	
➢ Power requirement: 3 phase 440 volt, 30 kVa.	
[B] Motor Driven	
➢ All position motor operated table from $+90^0$ to -12^0 trendelenburg position.	
➢ Automatic stop arrangement for manual operation in case of power failure.	
➢ Table has arrangement for manual operation in case of power failure.	
[C] Bucky	
➢ The bucky has counter balanced movement throughout the length of the table with 171/4" x 181/4" x-ray grid of 8:1 ratio and 103 lines.	
➢ It has stainless steel cassette tray to accommodate all sizes of cassettes.	
[D] Spot Film Device	
29. The spot film device mounted on table.	
30. It consists of 14" x 14" fluoroscopy lead glass and screen	
31. Grid ration of 6:1, 85 lines per inch, cough guard and stray radiation guard.	
32. It should be capable of taking 4 spot on 8" x 10" and 2 spot on 10" x 12" cassettes.	
[E] Column Stand	

Floor to ceiling column stand with counter balance vertical movement	
It can move horizontally on floor rails	
It can rotate 360^0	
It can be locked for various movements.	

(INST) (PART-III)
PART III: TECHNICAL BID
(IN DUPLICATE)

Name of Item: **Arthroscopic Light Source**		**Tender No: T 20**
Manufacturer:	Brand:	Model
Bidder Name:		**E.M.D.: Rs.3000/-**
Required Specifications:		Specifications available in offered model
➢ Lamp – Xenon Bulb 170 to 180 W Halogen Bulb 220 to 250 Watts		
➢ Power Supply – 100-120 / 200 – 240 VAC 50 -60 Hz		
➢ Power Supply Cord 400A		
➢ Light Intensity Manually adjustable		
➢ Light Cable fiber optic 1.5 to 2 meters		
➢ Light Outlets 1 or 2		
➢ Spare bulbs		
➢ Service contracts		
➢ Hospital Demonstration		
➢ CE certification		

(INST) (PART-III)
PART III: TECHNICAL BID
(IN DUPLICATE)

Name of Item: **Short Wave Diathermy (400 w)**		**Tender No: T 21**
Manufacturer:	Brand:	Model
Bidder Name:		**E.M.D.: Rs.1000/-**
Required Specifications:		Specifications available in offered model
➢ 400 wts Frequency 27.12 MHZ neon tester, time switch disc electrode, pad electrode sinus electrode.		

(INST) (PART-III)
PART III: TECHNICAL BID
(IN DUPLICATE)

Name of Item: **Continuous Passive Motion (CPM)**		**Tender No: T 22**
Manufacturer:	Brand:	Model
Bidder Name:		**E.M.D.: Rs.1000/-**
Required Specifications:		Specifications available in offered model
➢ Extension and Flexion indicator at knee position.		
➢ Treatment time 0 to 99 minutes with alarm.		
➢ Rest time 1 to 5 seconds.		
➢ Angle: - Extension 0 to 70 degree		
➢ Speed control adjustment		
➢ Patient safety switch for stopping and reversing the motion		

(INST) (PART-III)
PART III: TECHNICAL BID
(IN DUPLICATE)

Name of Item: **Stimulator or Combination Therapy Unit**		**Tender No: T 23**
Manufacturer:	Brand:	Model
Bidder Name:		**E.M.D.: Rs.1000/-**
Required Specifications:		Specifications available in offered model
➢ Plain ripple free D.C.current with variable amplitude		
➢ Constant current pulses of controllable amplitude and duration are precisely generated at.		
➢ 01, 03, 0.1, 0.3,1,3,10,100,300, ms at one pulse per second. Facilitate to plot strength duration can.		
➢ Faradic section is also built in with plain and surged faradic out put which makes a very useful instrument for diagnoses as well as for therapy.		
Extra feature For Combination Therapy Unit.		
Multiple waveforms on board:		
➢ 2 Pole IFCT, 4 Pole IFCT, TENS, Voltage Muscle Stimulator, Russian, Galvanic Current,		
➢ Diadynamic, High voltage Pulse Current, Micro current, Trabert, Monophonic, Rectangular, Monophasic Triangular Ultra-Sound Therapy-1 and 3 MH, pulse and		
Continuous Ultra-Sound Therapy:		

➢ Pulse and continuous ultrasound therapy (10%, 20%, 50%, & 100%).	
➢ Variable duty cycle outputs of16Hz.48Hz or 100Hz for ultrasound.	
➢ Variety of Ultrasound Sizes (Optimal)	
➢ Ultrasound heads are interchangble due to Electronic Signature system.	
➢ Head Warming feature and head contact control with visual and auditory feedback on head.	

(INST) (PART-III)
PART III: TECHNICAL BID
(IN DUPLICATE)

Name of Item: **Scanner Laser**		**Tender No: T 24**
Manufacturer:	**Brand:**	**Model**
Bidder Name:		**E.M.D.: Rs.3000/-**
Required Specifications:		Specifications available in offered model
➢ Visible Red Laser at 655nm wave length with output power 24 mw		
➢ Infrared Red Laser at 830 mn wave length with output power 120mw		
➢ Pulsed and Continuous output		
➢ Meridian point detector		
➢ Scanning mode: - Very useful for treating larger areas such as shoulder and lumber region without strain for the operator		
➢ Computerized LCD screen with control gives messages / information to the user on various Adjustments		
➢ Automatic electronic treatment timer with alarms		
➢ key operated switch prevent unauthorized operation		
➢ Mounted on stand and wheels for easy mobility		
➢ Supplied with protective goggles for the operative as well as for the patient.		

(INST) (PART-III)
PART III: TECHNICAL BID
(IN DUPLICATE)

Name of Item: **EMG/NCV/ Evoked Potential windows based system**	**Tender No: T 25**
Manufacturer: Brand: Model	
Bidder Name:	**E.M.D.: Rs.3000/-**
Required Specifications:	Specifications available in offered model
➢ Computerized & EMG/NCV/EP Systems.	
➢ Directly connect to USB, no external power supply required.	
➢ No Battery	
➢ Single screen surface EMG	
➢ Triggered EMG	
➢ Motor Nerve conduction studies	
➢ Sensory Nerve Conduction Studies.	
➢ F.wave /H-Relax / B link Relax / Repetitive stimulation	
➢ Soma to sensory Evoked potential	
➢ Auditory Evoked potential	
➢ Visual Evoked	

{Signature & Stamp of Bidder}

➢ Sensitivity -0.1-2000 mu	
➢ High cot-100,200,500 Hz/ 1,2,3,5,10 KHz	
➢ Impedance------ 100m ohm	
➢ CMRR----- 100 db input	
➢ low cot------- up to 500 Hz/1 kHz	
➢ Sweep speed ------- 1 to 500 ms/ div	
➢ Average ---------- 1 to 4,00,000	
➢ Stimulus intensity display, potarity reversal	
➢ Repetition rate -------- 0.5 Hz to 30pps in 10 steps.	
➢ Pulse width ---------- 0.5/0.1/0.2/0.3/0.5/0.8/1 ms	
➢ Tone Frequency ----------Left, Right, Binqural	
➢ Presentation---------- 0 to 130 db	
➢ Checker Board ---------4/8/16/32/64/ selectable pattern size	
➢ Marking of Latency / amplitude automatically or user	
➢ Foot switch par- Independent to allow hand free triggering single shot triggering & multiple shot triggering and shot	
➢ Stimulator for have Digital display / Stimulation swathe & store swathe	
➢ Trolley	
Computer	
➢ P-4	
➢ 256 DD & Ram	
➢ 80 GB hard disk	

➢ Combo Drive	
➢ 17” monitor	
➢ Key Board / Mouse	
Printer colored Laser	
➢ Micro controller based, 4 Line, 20 characters LCD Display, air regulator, moisture free compressor, printer port D.M.printer.	

(INST) (PART-III)
PART III: TECHNICAL BID
(IN DUPLICATE)

Name of Item: **Co2 Laser With Pattern Generator**		**Tender No: T 26**
Manufacturer:	Brand:	Model
Bidder Name:		**E.M.D.: Rs.10000/-**
Required Specifications		Specifications available in offered model
Co2 Laser with pattern generator facility		

INST) (PART-III)
PART III: TECHNICAL BID
(IN DUPLICATE)

Name of Item: **Skin Bank Refrigerator**	**Tender No: T 27**
Manufacturer: Brand: Model:	
Bidder Name:	**E.M.D.: Rs.3000/-**
Required Specifications:	Specifications available in offered model
Deep-freeze with maintenance of temperature of -40 degree centigrade.	
Adjustable temperature module with digital temperature clock.	
Two separated chambers, if possible.	

{Signature & Stamp of Bidder}

(INST) (PART-III)
PART III: TECHNICAL BID
(IN DUPLICATE)

Name of Item: **Microsurgical Instruments [Ergonomic Design]**	**Tender No: T 28**
Manufacturer: Brand: Model:	
Bidder Name:	**E.M.D.: Rs.3000/-**
Required Specifications:	Specifications available in offered model
➢Microsurgical Scissors	
• 5" 2 pc. [One curved and one straight]	
• 6" 2 pc. [One curved and one straight]	
• 7" 2 pc. [One curved and one straight]	
➢Microsurgical Forceps	
• 5" 2 pc. [One tooth and one non-toothed]	
• 6" 2 pc. [One tooth and one non-toothed]	
• 7" 2 pc. [One tooth and one non-toothed]	
➢Vessel dilators	
• 5" 2 pc. [One curved and one straight]	
• 6" 2 pc. [One curved and one straight]	
• 7" 2 pc. [One curved and one straight]	
➢Microsurgical Needle holders	

• 5" 2 pc. [One curved and one straight]	
• 6" 2 pc. [One curved and one straight]	
• 7" 2 pc. [One curved and one straight]	
➢Vascular clamps with controlled pressure (S & T - Imported)	
• 2A, 2V, 3A, 3V, 4A, 4V: two pairs each one coupled and one pair as singles	
➢Clamp Applicator	
• 2	

(INST) (PART-III)
PART III: TECHNICAL BID
(IN DUPLICATE)

Name of Item: **Surgical Operating Microscope For Plastic Surgery**		**Tender No: T 29**
Manufacturer:	Brand:	Model:
Bidder Name:		**E.M.D.: Rs.25000/-**
Required Specifications		Specifications available in offered model
➢ OPMI	1	
➢ Dust cover	1	
➢ S-light guide 2.0m	1	
➢ S7 floor stand, basic equipment	1	
➢ Electrical components kit for s7 stand	1	
➢ Standard set of panels-floor stand	1	
➢ Halogen illumination-s7	1	
➢ 180* inclinable binocular tube F=170 mm	1	
➢ Sleeves for tubes with screw thread	2	
➢ Push in wide field eye piece 12.5 x	2	
➢ Sterile drape for OPMI type 73 pack of 10	1	
➢ Hand grip for OPMI movena side handle	1	
➢ Pair of sterilizable handgrip cover	1	
Accessories		
Stereo Co-Observation System Consisting of		
➢ Stereo co-observation tube with two joints	1	
➢ Inclined binocular tube F=170mm	1	
➢ Sleeves for tubes with screw thread	2	
➢ Push in wide field eyepiece 12.5 x	2	

➢ Beam splitter 50 1	
CCTV Attachment Consisting of	
Beam splitter 50 1	
Video lens F=85 mm/c mount 1	
Medilive 1 CCD color video camera.	
C-mount	
Technical Specifications:	
➢ Magnification: 1:6 ratio in 5 steps	
➢ Focusing System: internal, manual, and continuously adjustable	
➢ Working Range: 200-415mm	
➢ Main Binocular Tube: straight tube or optional 0- 180* inclinable tube with 12.5 x or 10 x wide field eyepieces	
S7 Floor Stand & Ceiling Mount	
➢ Voltage: 115v / 230v+ _10A	
➢ Frequency: 50-60 Hz	
➢ Power Consumption: 115v, max, 10A	
➢ Weight: approx 170kg (with opmi on floor stand)	
➢ Approx 7kg (with opmi on ceiling mount with fixed height column)	
➢ Approx 125kg (with opmi on ceiling mount with power Column)	
➢ Compliance: DIN EN ISO 9001	
o EN 46001	
o ISO 13485	
Light Sources:	
Halogen Illumination	
➢ Main & back up illumination	
➢ 100w halogen lamp via light guide, backup lamp in automatic, fast action lamp changer	
Super Lux (Option)	
➢ Main & back up illumination	
➢ 180w xenon lamp with day light color temperature via light guide, back up xenon lamp in fast action lamp changer	

(INST) (PART-IV)
PART IV: COMMERCIAL BID
(IN DUPLICATE)

Tender No. & Name: _____
E.M.D. Rs:
Name of Bidder: _____

(TO BE SUBMITTED IN SEPARATE ENVELOPE MARKED "COMMERCIAL BID")
(DO NOT SUBMIT THIS PART IN THE SAME ENVELOP WITH PARTS I/II/III)

Sr. No.	Code No.	Name of Item	Unit	Mfgd. By	Basic rate	Tax GST/CST %	Excise Duty %	Freight/Insurance, installation charges etc.	Other duty, taxes & charges (pl.specify)	Final rate	Remarks
1.	2.	3.	4.	5.	6.	7.	8	9	10	11	12.

1. **The above commercial bid must be submitted in duplicate and bidder has to mention original copy and duplicate copy very clearly.**
2. Bidder should specify if GST/CST % mentioned in Column 7.
3. The above quoted rates are inclusive of Excise duty, V.A.T., Packing, Forwarding and insurance charges F.O.R. Destination at Central medical stores and Ahmedabad municipal corporation hospitals.
4. If more items are quoted than attach additional photocopy sheets.
5. Rates must be quoted per unit.
6. Rates should be quoted in Indian currency even though the item is imported.

{Signature & Stamp of Bidder}